

## APM-T 3053/2 Termo abrasion-Test machine



### Termo abrasion Test machine

With this test the chalking of outsole materials can be determined.

During the test a specimen which is taken out of the outsole is put on a rotating light - colored PVC-standard flooring (e.g. "Royal", color: 420-54, manufacturer: DLW) with a defined force. The staining of the flooring is determined according to the grey scale DIN EN 20105/A03.

#### Technical Data

Footprint: T/B/H 600 mm / 900 mm (+ Control unit) / 500 mm

Supply voltage: 200 - 230 V

Supply frequency: 50 - 60 Hz

Air pressure: 6bar

Reference N°: [APM-T 3053/2](#),

Phone: +49 6331 2490 0, E-Mail: [peter.schultheis@pfi-germany.com](mailto:peter.schultheis@pfi-germany.com)