

SE 3065/2

Ramp Tester



Ramp Tester

For assessing slip resistance of footwear and flooring materials conforming to DIN 4838 T 100, DIN 51097, DIN 51098 and DIN 51130.

For prototype testing of safety, protective and occupational footwear, determination of anti-slip properties (DIN 4843-100) is mandatory. A test person wearing the prototypes walks on a ramp of adjustable inclination, which can be fitted with different floor coverings. Inclusive of steel plate on carrier plate. Rest floor coverings not in the extent of supply. To prevent injuries, the test person wears safety belts.

Technical Data

Footprint: approx. 1500 × 2500 mm

Supply voltage: 230V

Supply frequency: 50 Hz

Reference N° **SE 3065/2**

Phone: +49 6331 2490 0, E-Mail: peter.schultheis@pfi-germany.com